



Declaration of Conformity

This European Declaration of Conformity is issued under the sole responsibility of the manufacturer.

MANUFACTURER			
Name of Company		Address	SRN
Bio/Data Corporation		155 Gibraltar Road, Horsham, PA 19044 U.S.A.	US-MF-000026991
AUTHORIZED REPRESENTATIVE			
Name of Company	Address	SRN	Telephone/email
Mdi Europa	Langenhagener Str. 71 30855 Langenhagen GERMANY	DE-AR-000006218	+49-511-3908 9531 – phone info@mdi-europa.com
PRODUCT IDENTIFICATION			
Product / Trade Name	Product Code / Catalog Number	Basic UDI-DI	
PDQ Platelet Function Centrifuge	106842	++G0561068425A	
Intended Purpose			
See Instructions for Use			
The PDQ Model Platelet Function Centrifuge is for in vitro diagnostic use for rapid separation of whole blood contained in original collection tubes.			
IVDR RISK CLASS / COMMON SPECIFICATIONS			
Device Classification		Common Specifications	
Class	A non-sterile	No relevant common specifications have been published yet.	
Rule	5c per Annex VIII of IVDR 2017/746		

Bio/Data Corporation, the Manufacturer, declares that the above-mentioned product meets the provision of the following EU legislation:

In Vitro Diagnostic Medical Device Regulation (IVDR) (EU) 2017/746
RoHS Directive 2015/863/EU

Conformity Statement:


Bio/Data Corporation confirms that the device covered by this declaration is in conformity with the (EU) IVDR 2017/746 Regulation and, if applicable, with any other relevant Union Legislation that provides for the issuing of an EU Declaration of Conformity.

Notified Body (NB) Identification, conformity assessment procedure performed and identification if the certificate or certificates issued:

Not applicable. The conformity assessment procedure for Class A Devices should be carried out, as a general rule, under the sole responsibility of manufacturers, since such devices pose a low risk to patients.
(Self-Declaration)

COMPANY REPRESENTATIVE: William M. Trolie

SIGNATURE: _____

A handwritten signature in black ink, appearing to read 'W. Trolie', written over a horizontal line.

TITLE/FUNCTION: Director of Quality Assurance & Regulatory Affairs for Bio/Data Corporation

PLACE: Bio/Data Corporation, Horsham, PA. 19044 U.S.A.

DATE: 9 March 2023