

PRODUCT DESCRIPTION

PAR/PAK® II is a Platelet Aggregation Combination Kit containing ADP (Adenosine-5'-Diphosphate), Collagen (Soluble Calf Skin, Type 1), and Epinephrine (Adrenaline) Reagents.

ADP Reagent is a lyophilized preparation of Adenosine-5'-Diphosphate. It is an essential component in platelet aggregation. ADP acts as an agonist or activator, binding to platelet receptors and triggering a series of biochemical events that lead to platelet activation and aggregation.

Collagen Reagent is a lyophilized preparation of Soluble Calf Skin (Type 1). Collagen Reagent induces platelet shape change and activates platelets. The activated platelets then release thrombotic compounds from their granules, which serve to recruit additional platelets to an injury site.

Epinephrine Reagent is a stabilized and lyophilized preparation of L-Adrenaline that activates the GP IIa adreno receptor causing platelet aggregation without shape change. Although it can enhance the response of platelets to other agonists, Epinephrine Reagent is a weak (reversible) agonist. It may or may not elicit a response in healthy people.

PAR/PAK II Combo Kit has been optimized for use with Light Transmission Aggregometers. It may also be used with other turbidometric or impedance analyzers, and flow cytometers.

INTENDED PURPOSE

PAR/PAK II Combo Kit is a convenience kit containing a combination of routine platelet aggregation reagents used to elicit aggregation and / or agglutination responses in Platelet Rich Plasma (PRP). The Kit includes ADP, Collagen, and Epinephrine Reagents.

DETECTION / MEASUREMENT

PAR/PAK II Combo Kit Reagents are used, in conjunction with other diluents and control samples, to measure changes of the light transmission in a Platelet Rich Plasma (PRP) test sample.

PRODUCT FUNCTION

PAR/PAK II Combo Kit provides insight into different aspects of platelet function / quality. This Kit aids in accessing various acquired and inherited platelet disorders or the efficacy of anti-platelet therapies.

SPECIFIC INFORMATION PROVIDED

PAR/PAK II Combo Kit Reagents are not intended for the detection of a specific disorder, condition, or risk factor.

ADP Reagent plays a pivotal role in platelet activation and aggregation. When ADP binds to specific receptors on the platelet surface, such as P2Y1 and P2Y12, it initiates intracellular signaling cascades. This activation induces rapid changes in platelet shape and the release of calcium ions through P2Y1 receptors, while P2Y12 activation sustains the response, ensuring stable aggregation. ADP Reagent is utilized to stimulate platelet activation and aggregation precisely by interacting with these ADP receptors. By observing platelet aggregation in response to ADP, clinicians can assess platelet function / quality related to abnormalities in platelet activation and aggregation. This process is crucial for understanding clot formation dynamics and evaluating the efficacy of anti-platelet therapies in preventing thrombotic events. ADP prompts the release of secondary mediators like Thromboxane A₂ (TXA₂), further amplifying platelet activation and aggregation.

Collagen Reagent initiates platelet activation and aggregation. Upon binding to glycoprotein receptors on the platelet surface, particularly glycoprotein VI (GP VI), Collagen sets off intracellular signaling cascades. This triggers rapid changes in platelet shape and the release of calcium ions through GP VI receptors, with sustained activation facilitated by integrin α2β1, ensuring stable aggregation. Utilized to precisely stimulate platelet activation and aggregation, Collagen Reagent interacts with these receptors, providing a means for clinicians to assess platelet function / quality and disorders linked to collagen-induced platelet activation abnormalities. This process is vital for comprehending clot formation dynamics and evaluating the efficacy of anti-platelet therapies inhibiting thrombotic events. Collagen prompts the release of secondary mediators, further amplifying platelet activation and aggregation.

Epinephrine Reagent plays a pivotal role in platelet activation and aggregation. Upon binding to specific receptors on the platelet surface, particularly α2-adrenergic receptors, epinephrine initiates intracellular signaling cascades. This cascade induces rapid changes in platelet shape and triggers the release of calcium ions, crucially mediated through α2-adrenergic receptor activation. The sustained response, essential for stable aggregation, is facilitated by α2-adrenergic receptor activation. Epinephrine

Reagent is instrumental in precisely stimulating platelet activation and aggregation by interacting with these adrenergic receptors. Observing platelet aggregation in response to Epinephrine Reagent allows clinicians to assess and evaluate platelet function / quality and disorders associated with abnormalities in platelet activation and aggregation. This process is pivotal for comprehending clot formation dynamics and evaluating the effectiveness of anti-platelet therapies in preventing thrombotic events. Epinephrine prompts the release of secondary mediators, further amplifying platelet activation and aggregation.

AUTOMATION

PAR/PAK II Combo Kit Reagents are intended for use in semi-automated and automated Light Transmission Platelet Aggregometers. These reagents may also be used with other turbidometric or impedance analyzers, and flow cytometers.

QUALITY / QUANTITY

There are no primary standards for the PAR/PAK II Combo Kit Reagents. The responses to these reagents are concentration dependent. A known normal donor should be tested with each new lot of PAR/PAK II Combo Kit Reagents. Standards organizations classify ADP, Collagen, and Epinephrine induced platelet aggregation as semi-quantitative or semi-qualitative.

PAR/PAK II Combo Kit comes packaged as 2 x 0.5 mL vials of ADP Reagent, 2 x 0.5 mL vials of Collagen Reagent, and 2 x 0.5 mL vials of Epinephrine Reagent. The working concentration of ADP is 200 μM, Collagen is 1.9 mg / mL, and Epinephrine is 100 μM.

SPECIMEN TYPE

The test specimen is prepared from sodium citrate anti-coagulated whole blood. The test sample is Platelet Rich Plasma (PRP). The test blank is Platelet Poor Plasma (PPP).

ADP, Collagen, and Epinephrine Reagents may be used with human or animal Platelet Rich Plasma (PRP) for routine platelet aggregation tests. Results are based on the concentration, extent, and rate of aggregation compared to a Platelet Poor Plasma (PPP) blank.

TESTING POPULATION

- Human: For ADP and Collagen the prevalence of platelet disorders is global and may vary by race, ethnicity, blood type, and other factors. The incidence is variable. For Epinephrine, the prevalence of abnormal Epinephrine Reagent aggregation is 16 – 20% in healthy people. It is global and may vary by race, ethnicity, blood type, and other factors. The incidence is variable.
- Anti-Platelet Drugs: For ADP, the prevalence and incidence are variable. 4% of the population over the age of 40 take Anti-Platelet Drugs, other than Aspirin. 33% (For Adults > 40); 16% Dual Anti-Platelet Therapy (DAPT); and 8% Anti-Platelet Therapy (APT). For Collagen, the prevalence of abnormal Collagen Reagent aggregation, contingent on estimated Aspirin usage, reaches up to one third of the population. Both Clopidogrel and the combination of Clopidogrel with Aspirin can influence Collagen-induced platelet aggregation. The incidence is variable. For Epinephrine, the prevalence and incidence are variable. The varying response rates to Epinephrine have been noted across different populations. Studies have demonstrated that Dual Anti-Platelet Therapy and Aspirin can influence Epinephrine-induced platelet aggregation.
- Inherited Platelet Disorders: For ADP, the prevalence and incidence are variable. There are 60 types; 75 Known Genes; Frequency 5/1000; Estimated 1-2% of the population. For Collagen, the prevalence and incidence are variable. There are 60 types of inherited platelet disorders that affect approximately 0.3% of the population. Certain inherited platelet defects, such as Glanzmann's Thrombasthenia and Storage Pool Disease, show no response to Collagen Reagents. For Epinephrine, the prevalence of abnormal epinephrine response in people varies with the defect. The incidence is variable.
- Animal: For ADP, Collagen, and Epinephrine, the prevalence and incidence are species dependent.

IN VITRO DIAGNOSTIC

PAR/PAK II Combo Kit contents are in vitro diagnostic reagents intended for Professional Laboratory Use Only. These Reagents are not intended for injection or ingestion.

INTENDED USER

PAR/PAK II Combo Kit Reagents are intended for Professional Laboratory Use by qualified personnel.

TEST PRINCIPLE

When introduced to a stirred, 37°C Platelet Rich Plasma (PRP) test sample, exogenous Reagents such as ADP, Collagen, and Epinephrine stimulate platelets, prompting them to undergo shape change and aggregate. This initial aggregation is called primary aggregation and is reversible. However, normal platelets possess the ability to release endogenous ADP from their granules, leading to a secondary, irreversible wave of aggregation. The Light Transmission Platelet Aggregometer effectively captures these changes by displaying parameters such as lag phase, shape change, and the rate and extent of aggregation over a predetermined testing period.

For Epinephrine, hyper-reactivity may be demonstrated. If so, the Sticky Platelet Procedure should be followed for confirmation. Not all healthy people will respond to Epinephrine Reagent.

CALIBRATORS AND CONTROLS

There are no calibrators or controls required for the PAR/PAK II Combo Kit. A known donor sample should be tested with each lot of ADP, Collagen, and Epinephrine Reagents. Responses are concentration dependent.

REAGENT LIMITATIONS

PAR/PAK II Combo Kit Reagents will perform as specified when the Instructions for Use are followed. The reagents must be used prior to the expiration date printed on each vial.

REAGENTS PROVIDED

REF	107650: 2 vials of ADP Reagent (0.5 mL)
	2 vials of Collagen Reagent (0.5 mL)
	2 vials of Epinephrine Reagent (0.5 mL)

REAGENTS AND MATERIALS REQUIRED BUT NOT PROVIDED

- Purified Water (Distilled, Deionized, Reagent Grade), pH 5.3 – 7.2 for reconstitution
- TRIS Buffered Saline (TBS) or 0.85% physiological saline for dilutions

 NOTE: USING BLOOD BANK SALINE WILL CAUSE ERRONEOUS RESULTS.

MATERIALS AND ACCESSORIES

- Platelet Aggregometer (Follow the Manufacturer's Instructions for Use)
- Centrifuge
- Electronic Pipette
- Pipette Tips 
- Aggregometer Test Tubes (Siliconized) 
- Aggregometer Stir Bars (Plastic Coated) 
- Plastic Sample Tubes and Caps (for Dilutions) 

 NOTE: DISPOSABLE ITEMS SUCH AS TEST TUBES, STIR BARS, SAMPLE TUBES, AND CAPS ARE FOR ONE TIME USE ONLY

STORAGE AND STABILITY

 ADP, Collagen, and Epinephrine, Reagents do not require temperature protection during shipment.

 Upon receipt, store ADP, Collagen, and Epinephrine Reagents at 2 – 8° C in their original packaging.

 Reconstituted ADP, Collagen, and Epinephrine Reagents are stable for 30 days when stored in their tightly capped, original containers at 2 – 8° C.

STERILITY

 PAR/PAK II Combo Kit Reagents are not sterile products. Be careful not to contaminate the product when pipetting the reconstituted or aliquoted reagents.

WARNINGS AND PRECAUTIONS

 Wear PPE in accordance with laboratory policies and practices when handling ADP, Collagen, and Epinephrine Reagents.

 Follow standard precautions when preparing test specimens and samples.

 Handle ADP, Collagen, and Epinephrine Reagents with care to avoid contamination during use.

 Avoid reagent evaporation by limiting air – liquid exchange surfaces.

 To ensure optimum test results, a known donor control sample should be run consecutively, without interruption.

 To preserve reagent stability, store remaining reagents in their tightly capped, original containers.

 Dispose of post-test materials in accordance with applicable regulations and laboratory policies.

 NOTE TO USER: ANY SERIOUS INCIDENT THAT OCCURS IN RELATION TO THIS PRODUCT SHALL BE REPORTED TO THE MANUFACTURER AND THE COMPETENT AUTHORITY OF THE MEMBER STATE IN WHICH THE USER AND / OR PATIENT ARE ESTABLISHED.

INFECTIOUS MATERIAL STATUS

PAR/PAK II Combo Kit Reagents do not contain any infectious materials. Test specimens and samples must be considered infectious and should be handled as if capable of transmitting infection. After testing, test specimens and samples must be disposed of in compliance with applicable regulations and laboratory policies.

SPECIAL FACILITIES

PAR/PAK II Combo Kit Reagents do not require the use of special facilities within a laboratory environment.

PREPARATION FOR USE

 NOTE: PAR/PAK II COMBO KIT REAGENTS MUST BE AT ROOM TEMPERATURE (15 – 28° C) PRIOR TO RECONSTITUTION. STORED REAGENTS MUST BE BROUGHT TO ROOM TEMPERATURE PRIOR TO USE.

RECONSTITUTION

The working concentration of reconstituted ADP is 200 µM, Collagen is 1.9 mg / mL, and Epinephrine is 100 µM. All final concentrations are based on adding 25 µL of ADP, Collagen, and Epinephrine Reagents to a 225 µL Platelet Rich Plasma (PRP) test sample.

- Reconstitute ADP, Collagen, and Epinephrine Reagents with 0.5 mL of Purified Water.
- Invert gently to mix.

 NOTE: EPINEPHRINE MAY APPEAR CLOUDY BUT WILL BECOME CLEAR TO PALE YELLOW WITHIN A FEW MINUTES.

- Reconstituted ADP, Collagen, and Epinephrine Reagents should be kept capped prior to use.

DILUTIONS

For BIPHASIC AGGREGATION

To demonstrate biphasic ADP aggregation, the Platelet Rich Plasma (PRP) may be tested with various dilutions of the reagent. Further dilutions may be made to determine the threshold concentration. The threshold concentration is the lowest concentration that elicits a primary aggregation response.

 NOTE: FOR DILUTIONS, USE TRIS BUFFERED SALINE (TBS) OR 0.85% PHYSIOLOGICAL SALINE.

TABLE 1: ADP DILUTION CHART

ADP REAGENT	TRIS BUFFERED SALINE	WORKING CONCENTRATION	FINAL CONCENTRATION
————	————	200 µM	20 µM
125 µM	125 µM	100 µM	10 µM
62 µM	188 µM	50 µM	5 µM
25 µM	225 µM	20 µM	2 µM

PATIENT PREPARATION

Patients should refrain from taking aspirin or using aspirin-containing medications and products, as well as other medications, supplements, or energy drinks known to affect platelet function for 7 – 10 days prior to specimen collection. Ingestion of fatty foods, dairy products, and smoking should be avoided for 12 hours before specimen collection.

 NOTE: CONSULTATION WITH A PHYSICIAN IS REQUIRED PRIOR TO MAKING ANY MEDICATION CHANGES.

SPECIMEN COLLECTION

The specimen should be collected with care to avoid stasis, hemolysis, contamination by tissue fluid and exposure to glass. Specimens must be kept at room temperature. Release the tourniquet as soon as blood begins to flow into the collection device.



PRACTICE STANDARD PRECAUTIONS THROUGHOUT THE SPECIMEN COLLECTION, SAMPLE PREPARATION, AND ANALYTICAL PROCESSES. DISPOSE OF SHARPS AND BIOHAZARDOUS WASTE IN ACCORDANCE WITH APPLICABLE REGULATIONS AND LABORATORY POLICIES.

Evacuated Specimen Collection Technique

- Use a 21g or 23g winged needle collection set for specimen collection
- Draw blood into plastic evacuated specimen collection tubes containing 3.2% (0.11 M) sodium citrate anti-coagulant
- Gently mix the specimen collection tube 4 - 5 times by inversion
- Write collection time on the specimen label
- Maintain specimen collection tubes at room temperature
- Remix specimen collection tubes prior to centrifugation

Syringe Collection Technique

- Use a 21g or 23g winged needle collection set for the venipuncture
- Draw 9.0 mL of blood into a plastic syringe, avoiding excess suction
- Clamp the winged needle tubing and disconnect the syringe
- Immediately and gently dispense the blood specimen into a plastic (polypropylene) tube containing 1.0 mL of 0.11 M sodium citrate anti-coagulant. The blood to anticoagulant ratio is 9 parts blood to 1 part anti-coagulant
- Cap the plastic tube
- Gently mix the specimen collection tube 4 - 5 times by inversion
- Write collection time on the specimen label
- Maintain specimen collection tubes at room temperature
- Remix specimen collection tubes prior to centrifugation



NOTE: WHEN THE PATIENT'S HEMATOCRIT IS LESS THAN 30% OR GREATER THAN 55%, THE BLOOD TO ANTI-COAGULANT RATIO MUST BE ADJUSTED. BLUE TOP EVACUATED SPECIMEN COLLECTION TUBES MUST CONTAIN 3.2% (0.11 M) SODIUM CITRATE ANTICOAGULANT, WHICH IS THE RECOMMENDED CONCENTRATION FOR PLATELET FUNCTION STUDIES.

SAMPLE PREPARATION

Platelet Rich Plasma (PRP)

- Centrifuge the anti-coagulated blood at 150 x g for 10 minutes at room temperature
- Examine the plasma layer for red cells
- If red cells are present, re-centrifuge for an additional 5 minutes
- Use a Pipette to transfer the PRP to a plastic container labeled PRP
- Remove the PRP from a point just below the middle of the PRP volume for consistent platelet count (*THE TOP OF THE VOLUME HAS A LOWER PLATELET COUNT AND THE BOTTOM IS MORE CONCENTRATED*)
- Cap the container
- Allow the container to stand at room temperature

Platelet Poor Plasma (PPP)

- Centrifuge the remaining PRP blood specimen at 2500 x g for 20 minutes
- Use a Pipette to transfer the PPP to a plastic container labeled PPP
- Cap the container
- Allow the container to stand at room temperature

ASSAY PROCEDURE

Routine Aggregation Procedure



NOTE: THIS IS A GENERAL PROCEDURE. FOLLOW THE INSTRUCTIONS FOR USE PROVIDED BY THE MANUFACTURER OF THE AGGREGOMETER IN USE.

Prepare a Blank for Each Patient



NOTE: EACH PATIENT MUST HAVE THEIR OWN BLANK. ONE PATIENT'S BLANK CANNOT BE USED FOR ANY OTHER PATIENT. THE PATIENT'S BLANK MUST BE PREPARED FROM THE PATIENT'S PLATELET POOR PLASMA (PPP) SPECIMEN. IF THE SAME PATIENT IS BEING TESTED ON MULTIPLE TEST WELLS, THE SAME PATIENT'S BLANK MAY BE USED FOR THOSE TEST WELLS.

- Label a test tube with the letter "B", test well #, and patient ID to identify the Blank
- Pipette 250 μ L of Platelet Poor Plasma (PPP) into the test tube (*DO NOT ADD A STIR BAR*)
- Place Blank aside for later use
- Repeat the steps above for each patient

Prepare Samples

- Label one to eight new test tubes with each patient ID and test well #
- Place the labeled test tubes into the correct well # 1 - 8 of the stirred sample incubation wells
- Add a stir bar to each test tube
- Pipette 225 μ L of Platelet Rich Plasma (PRP) sample into each test tube in the stirred sample incubation wells (*MAKE SURE THERE ARE NO BUBBLES*)
- Select the on-screen timer for each stirred sample incubation well in use and the warming count down will start
- The samples will incubate at 37° C for the pre-set time
- Set the 100% baseline (Blank)

- Place the appropriate previously prepared patient's Blank test tube into test well # 1
- Select BLANK to activate the test well
- The BLANK button will change to START
- Repeat the steps above for each test well being used for testing

Begin Testing

- Once the countdown timer reaches 0:00, press the timer button to stop each stirred sample incubation well
- Transfer the test tube in the stirred sample incubation well # 1 to test well # 1
- Repeat the step above for each test well, making sure all test tubes remain with their corresponding well #'s during transfer
- Close the pipette guides
- Select START for test well # 1
- Pipette 25 μ L of reagent directly into the Platelet Rich Plasma (PRP) test tube in test well # 1 (*DO NOT ALLOW REAGENT TO RUN DOWN THE WALL OF THE TEST TUBE OR PERMIT PIPETTE TIP TO BREAK THE SURFACE OF THE SAMPLE*)
- Select INJECT for test well # 1
- Repeat the steps above for each test well being used for testing
- The test will now run for the pre-set time (*OTHER MANUFACTURER'S TEST PROCEDURES MAY SPECIFY DIFFERENT TIMES OR VOLUMES*)



NOTE: USE A KNOWN DONOR AS A CONTROL SAMPLE. EACH LABORATORY SHOULD ESTABLISH AND VALIDATE ITS OWN TEST PROTOCOL AND VERIFY THE RESULTING PERFORMANCE OF ITS TEST SYSTEM (REAGENTS, INSTRUMENT, AND TEST PROTOCOL).

QUALITY CONTROL

For platelet aggregation studies, a known donor should be tested in the same manner as the patient to ensure test system performance and consistency. A new control should be included with each test series, and preferably with each new reagent lot or after instrument maintenance. Each laboratory must define its acceptable ranges for its patient population and verify the expected performance of the test system.

RESULTS

The aggregation patterns for PAR/PAK II Combo Kit Reagents are depicted in Figures 1 through 6.

Typical aggregation patterns induced by ADP Reagent are illustrated in Figures 1 through 2. When ADP Reagent is used at a final concentration of 20 μ M, it induces a large single wave of aggregation in normal Platelet Rich Plasma (PRP). At lower concentrations, ranging from 2 μ M to 10 μ M, two distinct waves of aggregation may be observed. The primary wave is the immediate response to the exogenous ADP introduced by the reagent, while the secondary wave is due to the release of endogenous ADP from the storage pool of nucleotides within the platelets.

In some normal PRP samples, concentration-dependent disaggregation may be observed, indicating a variable response to different ADP concentrations. Spike marks in the figures indicate the points at which the reagent was added, providing clear reference points for the timing of reagent introduction and its effects on the aggregation process.

Typical aggregation patterns induced by Collagen Reagent are illustrated in Figures 3 and 4, providing a detailed representation of the reagent's effects on Platelet Rich Plasma (PRP). Following the addition of Collagen Reagent to PRP, an initial lag phase occurs during which no aggregation is observed. After this lag phase, normal platelets will exhibit a noticeable shape change. Following the shape change, a large, single wave of aggregation is observed, demonstrating the robust response of the platelets to Collagen Reagent.

Spike marks in the figures indicate the exact points at which the reagent was added, providing clear reference points for the timing of reagent introduction and its effects on the aggregation process.

Typical aggregation patterns induced by Epinephrine Reagent are depicted in Figures 5 and 6, offering a comprehensive view of its effects on Platelet Rich Plasma (PRP). When the Epinephrine Reagent is added to normal PRP, it induces a biphasic response characterized by two distinct waves of aggregation. The first wave represents the initial platelet response to the reagent, while the second wave is due to the release of additional platelet agonists from the granules within the platelets, further amplifying the aggregation process.

This biphasic response is a hallmark of healthy PRP samples, indicating normal platelet function. Conversely, abnormal Epinephrine aggregation is identified when the final aggregation is less than 30%, as shown in Figure 6. Such a reduced response may indicate platelet dysfunction or other hematological abnormalities, providing valuable diagnostic information.

Spike indicators in the figures mark the exact points at which the reagent is added, offering clear reference points for the timing of reagent introduction. These markers are essential for correlating the addition of Epinephrine Reagent with the observed aggregation patterns, allowing for precise analysis of its immediate effects on the aggregation process.

FIGURE 1: ADP NORMAL AGGREGATION




FIGURE 2: ADP ABNORMAL AGGREGATION




FIGURE 3: COLLAGEN NORMAL AGGREGATION




FIGURE 4: COLLAGEN ABNORMAL AGGREGATION




FIGURE 5: EPINEPHRINE NORMAL AGGREGATION




FIGURE 6: EPINEPHRINE ABNORMAL AGGREGATION




TABLE 2: ADP, COLLAGEN, AND EPINEPHRINE RESULTS OBSERVED IN PLATELET FUNCTION DEFECTS

DEFECT	ADP REAGENT	COLLAGEN REAGENT	EPINEPHRINE REAGENT
A SPIRIN-LIKE	↓ or N	↓	↓ or N
THROMBASTHENIA	↓ ↓	↓	↓ ↓
STORAGE POOL DISEASE	↓	↓	↓
VON WILLEBRAND SYNDROME	N	N	N
BERNARD-SOULIER SYNDROME	N	N	N

↓ = Reduced Aggregation Resulting From a Decrease or Absence of Secondary Wave
↓ ↓ = Reduced Aggregation Resulting From a Decrease or Absence of Primary and Secondary Wave
N = Normal Response

EXPECTED VALUES

Each laboratory should establish expected ranges for each reagent at various concentrations used to induce aggregation (Table 3).

LIMITATIONS

In Light Transmission Aggregometry, the presence of red blood cells in the PRP will cause the observed aggregation to be reduced. The presence of platelets in the PPP will cause final aggregation to be increased. Spurious results may occur if the PRP platelet count is less than 75,000 platelets / cumm. PRP platelet counts can only be performed using the hemocytometer method. Compromised samples must be rejected. If the results are abnormal, the test should be repeated on another occasion. Each laboratory must establish reference ranges tailored to the population it serves, and the specific reagent concentrations used.

ANALYTICAL PERFORMANCE

Platelet aggregation, induced by commonly used reagents like ADP, Collagen, and Epinephrine Reagents, is a non-linear test system. Responses are based on the

difference between the patient's Platelet Rich Plasma (PRP) and Platelet Poor Plasma (PPP) light transmission and therefore, results are unique to that patient. Certain parameters are more prone to non-linearity than others. These include lag phase, primary slope, secondary slope, biphasic response and disaggregation. The non-linearity is caused by many factors such as the reaction chemistry and instrumentation. Platelet aggregation displays the response rate or activity and does not quantify the reactants or their concentrations.

In platelet aggregation, accuracy is a relative parameter and is dependent on the test system. The limitations of platelet aggregation make it difficult to provide typical precision or reproducibility ranges.

The variability in linearity, precision and reproducibility of results in ADP, Collagen, and Epinephrine Reagent-based test systems is acknowledged by multiple standards organizations. The commonly accepted CV is $\pm 15\%$.

Test to Test Reproducibility:	less than $\pm 7.5\%$
Instrument to Instrument Reproducibility:	less than $\pm 15.0\%$
Reagent Lot to Lot Variability:	less than $\pm 10.5\%$
Laboratory to Laboratory (System to System)	less than $\pm 12.5\%$

TABLE 3: EXPECTED RESULTS FOR PLATELET AGGREGATION RESPONSES IN NORMAL DONORS

Total Aggregation at 6 Minutes

Parameter	Units	ADP REAGENT	COLLAGEN REAGENT	EPINEPHRINE REAGENT
Final Concentration		20.0 μM	0.19 mg / mL	10.0 μM
Primary Aggregation	%	81	85	87
Primary Slope		54	55	20
Secondary (Biphasic) Aggregation	%	Yes	No	Yes
Secondary Slope		Variable	0	Variable
Area Under The Curve	Minutes	320	524	540
Lag Phase	Seconds	< 10	< 60	0
Disaggregation	%	Yes	Yes	Yes
Maximum Aggregation	%	≥ 89	≥ 99	≥ 104
Final Aggregation	%	63 - 89	61 - 99	51 - 104

 NOTE: ADJUSTING PLATELET COUNTS IS NOT RECOMMENDED

REFERENCES

Allain JP, Cooper HA, Wagner RH, Brinkhous KM. Platelets fixed with paraformaldehyde: a new reagent for assay of von Willebrand factor and platelet aggregating factor. J Lab Clin Med. 1975 Feb;85(2):318-28.

Angiolillo DJ, Ueno M, Goto S. Basic principles of platelet biology and clinical implications. Circ J. 2010 Apr;74(4):597-607.

Born GV, Cross MJ. The Aggregation of Blood Platelets. J Physiol. 1963 Aug; 168(1):178-95.

Brinkhous KM, Read MS. Preservation of platelet receptors for platelet aggregating factor/von Willebrand factor by air drying, freezing, or lyophilization: new stable platelet preparations for von Willebrand factor assays. Thromb Res. 1978 Oct;13(4):591-7.

Bye A, Lewis Y, O'Grady J. Effect of a single oral dose of aspirin on the platelet aggregation response to arachidonic acid. Br J Clin Pharmacol. 1979 Mar; 7(3):283-6.

Cattaneo M, Cerletti C, Harrison P, Hayward CP, Kenny D, Nugent D, Nurden P, Rao AK, Schmaier AH, Watson SP, Lussana F, Puglione MT, Michelson AD. Recommendations for the Standardization of Light Transmission Aggregometry: A Consensus of the Working Party from the Platelet Physiology Subcommittee of SSC/ISTH. J Thromb Haemost. 2013 Apr 10.

CLSI. Procedures for the Handling and Processing of Blood Specimens for Common Laboratory Tests; Approved Guideline—Fourth Edition. CLSI document H18-A4. Wayne, PA: Clinical and Laboratory Standards Institute; 2010.

CLSI. Protection of Laboratory Workers from Occupationally Acquired Infections, Approved Guideline - Fourth Edition. CLSI document M29-A4. Wayne, PA: Clinical and Laboratory Standards Institute; 2010.

CLSI. Platelet Function Testing by Aggregometry, Approved Guideline - Fourth Edition. CLSI document H58-A. Wayne, PA: Clinical and Laboratory Standards Institute; 2008.

CLSI. Collection, Transport and Processing for Plasma Based Coagulation Assays and Molecular Hemostasis Assays, Approved Guideline - Fifth Edition. CLSI document H21-A5. Wayne, PA: Clinical and Laboratory Standards Institute; 2008.

CLSI. Clinical Laboratory Safety, Approved Guideline - Third Edition. CLSI document GP17-A3. Wayne, PA: Clinical and Laboratory Standards Institute; 2012.

Day HJ, Holmsen H. Laboratory tests of platelet function. Ann Clin Lab Sci (1971). 1972 Jan-Feb; 2(1):63-74.

Day HJ, Rao AK. Evaluation of platelet function. Semin Hematol. 1986 Apr;23(2):89-101.

Eichelberger, JW. Kinetic (Slope) Measurement of Platelet Aggregation. Bio/Data Corporation, Horsham, PA; 1984.

Favaloro EJ, Gosselin RC, Pasalic L, Lippi G. Post-analytical issues in hemostasis and thrombosis testing: An update. In EJF, RCG, editors, Hemostasis and Thrombosis: Methods and Protocols. 2nd ed. New York: Humana Press. 2023. p. 787-811. (Methods in Molecular Biology).

Federici AB, Lee CA, Berntorp EE, Lillicrap D, Montgomery RR. Von Willebrand

Disease: Basic and Clinical Aspects. 2011.

Garner JS. Guideline for isolation precautions in hospitals. The Hospital Infection Control Practices Advisory Committee. Infect Control Hosp Epidemiol. 1996 Jan;17(1):53-80.

Howard MA, Firkin BG. Ristocetin--a new tool in the investigation of platelet aggregation. Thromb Diath Haemorrh. 1971 Oct 31; 26(2): 362-9.

Israels SJ, El-Ekiaby M, Quiroga T, Mezzano D. Inherited disorders of platelet function and challenges to diagnosis of mucocutaneous bleeding. Haemophilia. 2010 Jul;16 Suppl 5:152-9.

Kambayashi J, Shinoki N, Nakamura T, Ariyoshi H, Kawasaki T, Sakon M, Monden M. Prevalence of impaired responsiveness to epinephrine in platelets among Japanese. Thromb Res. 1996 Jan 1;81(1):85-90.

Levine PH. The effect of thrombocytopenia on the determination of platelet aggregation. Am J Clin Pathol. 1976 Jan;65(1):79-82.

Linnemann B, Schwonberg J, Mani H, Prochnow S, Lindhoff-Last E. Standardization of light transmittance aggregometry for monitoring antiplatelet therapy: an adjustment for platelet count is not necessary. J Thromb Haemost. 2008 Apr;6(4):677-83.

Marcus AJ, Coleman RW, Hirsh J, Ivarer VJ, Salzman EW. Hemostasis and thrombosis: Basic Principles and Clinical Practice. Vol. 472. Philadelphia: JB Lippincott Company; 1982.

Michelson, AD. Platelets. Third Edition. Amsterdam: Academic Press; 2013.

Mills DC, Robb IA, Roberts GC. The release of nucleotides, 5-hydroxytryptamine and enzymes from human blood platelets during aggregation. J Physiol. 1968 Apr;195(3):715-29.

Moncada S, Vane JR. Arachidonic acid metabolites and the interactions between platelets and blood-vessel walls. N Engl J Med. 1979 May 17;300(20):1142-7.

NCCLS. Assays of von Willebrand Factor Antigen and Ristocetin Cofactor Activity; Approved Guideline. NCCLS document H51-A. NCCLS, 940 West Valley Road, Suite 1400, Wayne, Pennsylvania 19087-1898 USA, 2002.

O'Donnell CJ, Larson MG, Feng D, Sutherland PA, Lindpaintner K, Myers RH, D'Agostino RA, Levy D, Toeller GH; Framingham Heart Study. Genetic and environmental contributions to platelet aggregation: the Framingham heart study. Circulation. 2001 Jun 26;103(25):3051-6.

Owen CA Jr, Bowie EJW, Thompson JH Jr. The Diagnosis of Bleeding Disorders. 2nd ed. Little, Brown, and Company; 1975.

Palma-Barqueros V, Revilla N, Sánchez A, Zamora Cánovas A, Rodríguez-Alén A, Marín-Quilez A, González-Porras JR, Vicente V, Lozano ML, Bastida JM, Rivera J. Inherited Platelet Disorders: An Updated Overview. Int J Mol Sci. 2021 Apr 26;22(9):4521.

Siegel JD, Rhinehart E, Jackson M, Chiarello L; Health Care Infection Control Practices Advisory Committee. 2007 Guideline for Isolation Precautions: Preventing Transmission of Infectious Agents in Health Care Settings. Am J Infect Control. 2007 Dec;35(10 Suppl 2):S65-164.

The Hospital Infection Control Practices Advisory Committee, Centers for disease Control and Prevention, Public Health Service, U.S. Department of Health and Human Services. Guideline for isolation precautions in hospitals Part II. Recommendations for isolation precautions in hospitals. American Journal of Infection Control. 1996; Vol 24, Issue 1: 32-52.

Triplett DA, et al. Platelet function: laboratory evaluation and clinical application. Chicago, IL: American Society for Clinical Pathology 1978.

Weiss HJ. Aspirin and Platelets in Drugs and Hematologic Reactions. New York, NY: Dimittov and Nodine, eds. Grune and Stratton. 1974.

White, M.M., and Jennings, L.K. Platelet Protocols: Research and Clinical Laboratory Procedures, Academic Press, Inc.; 1999.

Williams WJ, Beutler E, Erslev AJ, Rundles RW. Hematology. New York, NY: McGraw-Hill. 1977.

SYMBOLS

	Bio-Hazardous
	Catalog Number
	Caution
	CE Marked & Registered Product
	Consult Instructions For Use
	European Union Representative
	In Vitro Diagnostic Device
	Manufacturer
	Must Read
	Non-Sterile
	Single Use Only
	Temperature Limitations
	United Kingdom Marked & Registered Product
	United Kingdom Representative

REVISION HISTORY

Document No: 101314 Revision: AA, 01/2025

- Modified Testing Instructions
- Implemented IVDR Regulatory Requirements
- Reformatted and Reconfigured to Enhance Operator Use

For a complete product catalog, please visit our website at www.biodatacorp.com or contact our Customer Service Department.

THE BIO/DATA CORPORATION PRODUCT LINE INCLUDES GENERAL PURPOSE, PROFESSIONAL LABORATORY USE REAGENTS INTENDED TO INDUCE AND REPORT PLATELET FUNCTION ACTIVITY AND RESPONSES. THIS PRODUCT IS WARRANTED TO PERFORM AS DESCRIBED IN ITS LABELING INCLUDING THE INSTRUCTIONS FOR USE. BIO/DATA CORPORATION MAKES NO CLAIM OR WARRANTY, EXPRESSED OR IMPLIED, OF THE CAPABILITY, FITNESS, OR MERCHANTABILITY FOR ANY OTHER PURPOSE. IN NO EVENT SHALL BIO/DATA CORPORATION BE LIABLE FOR ANY CONSEQUENTIAL DAMAGES ARISING OUT OF AFORESAID EXPRESSED WARRANTY.



155 Gibraltar Road
Horsham, PA 19044 USA

Worldwide: +1 215-441-4000
USA: 1-800-257-3282
FAX Worldwide: +1 215-443-8820
customer.service@biodatacorp.com

©BIO/DATA CORPORATION 2025



101310



**BIO/DATA
CORPORATION**

AN ISO 13485 REGISTERED COMPANY

www.biodatacorp.com

PROUDLY MANUFACTURED IN THE USA



mdi Europa GmbH
Langenhagener Str. 71
D-30855 Langenhagen GERMANY



Alpha Laboratories
40 Parham Drive Eastleigh
SO50 4NU Hampshire UNITED KINGDOM



PAR/PAK II INSTRUCTIONS FOR USE # 101314 REV AA