

PRODUCT DESCRIPTION

von Willebrand Factor Normal Control Plasma is lyophilized citrated human plasma which has been verified for a von Willebrand Factor activity level of 80 to 120 percent.

INTENDED USE

von Willebrand Factor Normal Control Plasma is prepared from a pool of infectious disease negative normal human plasma to be used in the same manner as a patient Platelet Poor Sample to verify the performance and sensitivity of the Ristocetin Cofactor Activity Assay.

PRINCIPLE

The von Willebrand Factor (ristocetin cofactor) is a property of the protein which is quantitated in plasma for the diagnosis and evaluation of von Willebrand syndrome.^{4,5} When assayed as a patient specimen, vW Normal Control Plasma ensures the sensitivity of the test system to deficiencies in the von Willebrand factor.

PRECAUTIONS

von Willebrand Factor Normal Control Plasma is for **PROFESSIONAL LABORATORY USE ONLY AND IN-VITRO DIAGNOSTIC USE ONLY AND NOT FOR INJECTION OR INGESTION**. The plasma has been tested at the source and found to be negative for HIV-1Ag, anti-HIV-1/2, Hepatitis B surface antigen, Hepatitis C antibody, and negative by a serological test for syphilis. However, all plasma of human origin should be handled as being potentially hazardous.

MATERIALS PROVIDED

Normal Control Plasma, 3 X 0.5mL. Store at 2° to 8° C prior to reconstitution.

MATERIALS REQUIRED BUT NOT PROVIDED

1. Platelet Aggregometer
2. Purified water (distilled, deionized or reagent grade), pH 5.3 - 7.2
3. Pipettors (0.5mL volumes)
4. Disposable Stir bars
5. Aggregometer cuvettes

INSTRUMENTATION

Normal Control Plasma will perform as described when used on most optical platelet aggregometers¹. Follow the manufacturer's instructions for operating the aggregometer in use.

RECONSTITUTION

NOTE: Reagents must be at room temperature (15° to 28°C) prior to reconstitution. Stored reagent must be brought to room temperature prior to use.

Reconstitute a vial of Normal Control Plasma with 0.5mL purified water.

REAGENT STORAGE

The reconstituted Normal Control Plasma is stable for 8 hours when stored at 2° - 8°C in its original tightly sealed container. Dilutions are considerably less stable and should be used within 40 minutes of preparation.

QUALITY CONTROL

Laboratories should follow generally accepted quality control practices when proficiency testing is not available.

To assure proper instrument operation and reagent performance, the use of vW Normal Control Plasma and a von Willebrand Factor Deficient (Abnormal Control) is recommended for each day tests are performed.

EXPECTED VALUES

Normal Control Plasma will yield von Willebrand factor assay results within the normal reference range, typically 80 to 120% activity (0.8 - 1.2 IU).⁶ The ability to generate a quantitative value in this range is dependent upon the sensitivity of the assay system in use.⁵ Since normal reference ranges for von Willebrand factor reported in the literature are dependent on blood type, each laboratory should establish blood type specific reference ranges for its patient population.⁷ Additionally each laboratory should establish its "expected" ranges for all control material used.

Unexpected assay values should be evaluated to determine the source of deviation. The following may contribute to erroneous results:

1. Non specific agglutination of reagent platelets
2. Incorrect ristocetin concentration
3. Improper reconstitution of normal control plasma
4. Instrument malfunction
5. Procedural error(s)

LIMITATIONS

The von Willebrand Factor Normal Control Plasma is provided to check and verify the performance of the Ristocetin Cofactor assay system including reagents and instrumentation. The use of this material is dependent upon the establishment of expected ranges for the total assay system and not specifically a recovery of a defined assay value. The user must establish the suitability of this material for their specific application and instrumentation.

PERFORMANCE CHARACTERISTICS

Studies have shown that this product will perform as described prior to its expiration date when procedural and storage directions are followed.

Linearity

The ristocetin cofactor assay is a non-linear test system. The non-linearity is caused by many factors such as the reaction chemistry and instrumentation. The ristocetin cofactor assay measures a response rate or activity that is not a quantitative measure of the reactants or their concentration.

ACCURACY, PRECISION AND REPRODUCIBILITY**Accuracy**

In the ristocetin cofactor assay, accuracy is a relative parameter and is dependent on the test system.

Precision and Reproducibility

The limitations of the ristocetin cofactor assay make it difficult to provide typical precision or reproducibility ranges. However, there is an experienced based consensus for these parameters (see below). Each laboratory must establish its own limits for test acceptability.

Test to Test Reproducibility:	less than ± 10%
Instrument to Instrument Reproducibility:	less than ± 15%
Reagent Lot to Lot Variation:	less than ± 10.5%
Laboratory to Laboratory (same test system):	less than ± 12.5%

REFERENCES

1. Born, GVR and Cross, MJ. The Aggregation of Blood Platelets. J. Physiol [London] 168:178, 1963.
2. Weiss HJ: Aspirin and platelets in drugs and hematologic reactions. Dimittov and Nodine (eds.). Grune and Stratton, New York, 1974.
3. Triplett DA, Harms CS, Newhouse P, Clark C: Platelet Function. Laboratory Evaluation and Clinical Application. ASCP, 1978.
4. Gralnck HR, Sultan Y, Collier BS: von Willebrand's disease: Combined qualitative and quantitative abnormalities. N Eng J Med 296(18):1024, 1977.
5. Bowie EJW, Owen CA: Abnormalities of factor VIII. In Triplett DA: Laboratory evaluation of coagulation. pg 116, American Society of Clinical Pathologists Press, Chicago, 1982.
6. Zuzel M, Nilsson IM, Aberg M: A method for measuring plasma ristocetin cofactor activity. Thromb Res 12(5): 745, 1978.
7. National Committee for Clinical Laboratory Standards, NCCLS. Assays of von Willebrand Factor Antigen and Ristocetin Cofactor Activity: Approved Guideline. HS1-A Vol 22 No 20. Sep'02.

For a complete list of available products please go to our web site www.biodatacorp.com or contact customer service below.

THE BIO/DATA CORPORATION PRODUCT LINE INCLUDES GENERAL PURPOSE, PROFESSIONAL LABORATORY USE REAGENTS INTENDED TO INDUCE AND REPORT PLATELET FUNCTION ACTIVITY AND RESPONSES. THIS PRODUCT IS WARRANTED TO PERFORM AS DESCRIBED IN ITS LABELING INCLUDING THE INSTRUCTIONS FOR USE. BIO/DATA CORPORATION MAKES NO CLAIM OR WARRANTY, EXPRESSED OR IMPLIED, OF THE CAPABILITY, FITNESS, OR MERCHANTABILITY FOR ANY OTHER PURPOSE. IN NO EVENT SHALL BIO/DATA CORPORATION BE LIABLE FOR ANY CONSEQUENTIAL DAMAGES ARISING OUT OF AFORESAID EXPRESSED WARRANTY.



155 Gibraltar Road, Horsham, PA 19044 U.S.A.
(800) 257-3282 U.S.A. (215) 441-4000 Worldwide
(215) 443-8820 Fax Worldwide
E-mail: customer.service@biodatacorp.com
Internet: www.biodatacorp.com
An ISO 13485 Registered Company



EMERGO EUROPE, Prinsessegracht 20, 2514 AP The Hague, The Netherlands