



## Declaration of Conformity

This European Declaration of Conformity is issued under the sole responsibility of the manufacturer.

| <b>MANUFACTURER</b>  |   |  |   |
|--|---|--|---|
| Name of Company  |   | Address  | SRN   |
| Bio/Data Corporation   |   | 155 Gibraltar Road,<br>Horsham, PA 19044 U.S.A.            | US-MF-00002691  |
| <b>AUTHORIZED REPRESENTATIVE</b>   |   |  |   |
| Name of Company  | Address   | SRN  | Telephone/email   |
| mdi Europa GmbH  | Langenhagener Str. 71<br>D-30855 Langenhagen<br>GERMANY | DE-AR-000006218  | +49-511-3908 9531 – phone<br><a href="mailto:info@mdi-europa.com">info@mdi-europa.com</a> |
| <b>PRODUCT IDENTIFICATION</b>  |   |  |   |
| Product / Trade Name   |   | Product Code / Catalog Number                              | Basic UDI-DI  |
| PDQ, Platelet Function Centrifuge  |   | 106842   | ++G0561068425A  |
| Intended Purpose   |   |  | Photo   |
| See Instructions for Use   |   |  | See website <a href="http://www.biodatacorp.com">www.biodatacorp.com</a>                  |
| The PDQ Platelet Function Centrifuge prepares samples for Platelet Aggregation Testing. The PDQ is for in vitro diagnostic use for rapid separation of whole blood contained in original collection tubes. The PDQ is a microprocessor controlled high-speed bench top centrifuge designed to rapidly separate blood in the original evacuated or syringe type collection tubes. |   |  |   |
| <b>IVDR RISK CLASS / COMMON SPECIFICATIONS</b>   |   |  |   |
| Device Classification  |   | Common Specifications                                      |   |
| Class  | A non-sterile   | No relevant common specifications have been published yet. |   |
| Rule   | 5c per Annex VIII of IVDR 2017/746                      |  |   |

155 Gibraltar Road, Horsham, PA 19044 U.S.A.  
 Worldwide: (215) 441-4000 U.S.A.: (800) 257-3282  
 Fax Worldwide: (215) 443-8820  
[www.biodatacorp.com](http://www.biodatacorp.com) e-mail: [customer.service@biodatacorp.com](mailto:customer.service@biodatacorp.com)  
 An ISO 13485 Registered Company



An ISO 13485 Registered Company

[www.biodatacorp.com](http://www.biodatacorp.com)  
[customer.service@biodatacorp.com](mailto:customer.service@biodatacorp.com)

**Bio/Data Corporation, the Manufacturer, declares that the above-mentioned product meets the provision of the following EU legislation:**

In Vitro Diagnostic Medical Device Regulation (IVDR) (EU) 2017/746  
RoHS Directive 2015/863/EU

**Conformity Statement:**

Bio/Data Corporation confirms that the device covered by this declaration is in conformity with the (EU) IVDR 2017/746 Regulation and, if applicable, with any other relevant Union Legislation that provides for the issuing of an EU Declaration of Conformity.

**Notified Body (NB) Identification, conformity assessment procedure performed and identification if the certificate or certificates issued:**

Not applicable. The conformity assessment procedure for Class A Devices should be carried out, as a general rule, under the sole responsibility of manufacturers, since such devices pose a low risk to patients.  
(Self-Declaration)

COMPANY REPRESENTATIVE: William M. Trolio

SIGNATURE: \_\_\_\_\_

A handwritten signature in blue ink, appearing to read "W. M. Trolio", is written over a horizontal line.

TITLE/FUNCTION: Director of Quality Assurance & Regulatory Affairs for Bio/Data Corporation

PLACE: Bio/Data Corporation, Horsham, PA. 19044 U.S.A.

DATE: 1 April 2023