



## Declaration of Conformity

This European Declaration of Conformity is issued under the sole responsibility of the manufacturer.

MANUFACTURER			
Name of Company		Address	SRN
Bio/Data Corporation		155 Gibraltar Road, Horsham, PA 19044 U.S.A.	US-MF-000026991
Name of Company	Address	SRN	Telephone/email
Mdi Europa	Langenhagener Str. 71 30855 Langenhagen GERMANY	DE-AR-000006218	+49-511-3908 9531 – phone <a href="mailto:info@mdi-europa.com">info@mdi-europa.com</a>
Product / Trade Name	Product Code / Catalog Number	Basic UDI-DI	
vW Abnormal Control Plasma 3 x 0.5mL	101270	++G0561012703E	
Intended Purpose		Photo	
See Instructions for Use		See website <a href="http://www.biodatacorp.com">www.biodatacorp.com</a>	
Abnormal Control Plasma is prepared from a pool of normal infection negative human plasma which has been partially depleted of vW Factor and then lyophilized. It is used to verify the performance and sensitivity of the Ristocetin CoFactor Activity test.			
Device Classification		Common Specifications	
Class	A non-sterile	No relevant common specifications have been published yet.	
Rule	5a per Annex VIII of IVDR 2017/746		

**Bio/Data Corporation, the Manufacturer, declares that the above-mentioned product meets the provision of the following EU legislation:**

In Vitro Diagnostic Medical Device Regulation (IVDR) (EU) 2017/746

**Conformity Statement:**

Bio/Data Corporation confirms that the device covered by this declaration is in conformity with the (EU) IVDR 2017/746 Regulation and, if applicable, with any other relevant Union Legislation that provides for the issuing of an EU Declaration of Conformity.

**Notified Body (NB) Identification, conformity assessment procedure performed and identification if the certificate or certificates issued:**

Not applicable. The conformity assessment procedure for Class A Devices should be carried out, as a general rule, under the sole responsibility of manufacturers, since such devices pose a low risk to patients. (Self-Declaration)

COMPANY REPRESENTATIVE: William M. Trolio

SIGNATURE: \_\_\_\_\_



TITLE/FUNCTION: Director of Quality Assurance & Regulatory Affairs for Bio/Data Corporation

PLACE: Bio/Data Corporation, Horsham, PA. 19044 U.S.A.

DATE: 3 February 2023