

Declaration of Conformity

This European Declaration of Conformity is issued under the sole responsibility of the manufacturer.

MANUF	ACTURER					
Name of Company		Address		SRN		
Bio/Data Corporation		155 Gibraltar Road,		US-MF-000026991		
		Horsham, PA 19044 U.S.A.				
AUTHOR	RIZED REPR	ESENT	ATIVE			
Name of	Name of Company Addre		ess	SRN	Telephone/email	
mdi Euro	pa GmbH	Lange	nhagener Str. 71	DE-AR-000006218	+49-511-3908 9531 – phone	
		D-30855 Langenhagen GERMANY			info@mdi-europa.com	
		GERIVIANT				
PRODUC						
IDENTIF	CATION					
Product / Trade Name		Product Code / Catalog Number		Basic UDI-DI		
Beta/Pak		101580		++G0561015803Y		
Intended Purpose					Photo	
See Instructions for Use					See website www.biodatacorp.com	
BETA/Pak (ADP, Collagen, and Ristocetin) is a convenience kit						
containing a combination of routine platelet aggregation reagents						
used to elicit responses in Platelet Rich Plasma as well as an						
agglutination response that may be induced by the Ristocetin reagent. BETA/Pak contains ADP, Collagen and Ristocetin.						
reagene. 521791 an contains 7151, conagen and histocetin.						
IVDR RISK CLASS / COMMON SPECIFICATIONS						
Device Classification			Common Specifications			
Class	A non-st	erile	No relevant common specifications have been published yet.			
Rule	•	5a per Annex				
	VIII of I\					
	2017/7	46				



www.biodatacorp.com customer.service@biodatacorp.com

Bio/Data Corporation, the Manufacturer, declares that the above-mentioned product meets the provision of the following EU legislation:

In Vitro Diagnostic Medical Device Regulation (IVDR) (EU) 2017/746

Conformity Statement:

Bio/Data Corporation confirms that the device covered by this declaration is in conformity with the (EU) IVDR 2017/746 Regulation and, if applicable, with any other relevant Union Legislation that provides for the issuing of an EU Declaration of Conformity.

Notified Body (NB) Identification, conformity assessment procedure performed and identification if the certificate or certificates issued:

Not applicable. The conformity assessment procedure for Class A Devices should be carried out, as a general rule, under the sole responsibility of manufacturers, since such devices pose a low risk to patients. (Self-Declaration)

COMPANY REPRESENTATIVE: William M. Trolio	SIGNATURE:
TITLE/FUNCTION: Director of Quality Assurance & Regu	ulatory Affairs for Bio/Data Corporation
, , , , , , , , , , , , , , , , , , , ,	, , , , , , , , , , , , , , , , , , , ,
PLACE: Bio/Data Corporation, Horsham, PA. 19044 U.	S A DATE: 1 April 2022
PLACE. Dio/Data Corporation, noisham, PA. 19044 O.	S.A. DATE : 1 April 2023