



EU Declaration of Conformity

This Declaration of Conformity is issued under the sole responsibility of the manufacturer.

| MANUFACTURER | | | |
|--|---|--|---|
| Name of Company | Address | SRN | |
| Bio/Data Corporation | 155 Gibraltar Road Horsham, PA 19044 USA | US-MF-000026991 | |
| AUTHORIZED REPRESENTATIVE | | | |
| Name of Company | Address | SRN | Telephone / Email |
| mdi Europa GmbH | Langenhagener Str. 71 D-30855 Langenhagen GERMANY | DE-AR-000006218 | +49-511-3908 9531 – phone info@mdi-europa.com |
| PRODUCT IDENTIFICATION | | | |
| Product / Trade Name | Product Code / Catalog Number | Basic UDI-DI | EMDN Code |
| vW Abnormal Control Plasma | 101270 | ++G0561012703E | W01030199 |
| Intended Purpose | | Photo | |
| See Instructions for Use | | See website www.biodatacorp.com | |
| vW Abnormal Control Plasma is prepared from a pool of normal infection negative human plasma which has been partially depleted of vW Factor and then lyophilized. It is used to verify the performance and sensitivity of the Ristocetin Cofactor Activity test. | | | |
| IVDR RISK CLASS / COMMON SPECIFICATIONS | | | |
| Device Classification | Common Specifications | | |
| Class A (non-sterile) | None currently published for this device type. | | |
| Rule 5(a) per Annex VIII of Regulation (EU) 2017/746 | | | |

An ISO 13485 Registered Company

155 Gibraltar Road, Horsham, PA 19044 USA

+1 215-441-4000 Worldwide / 1-800-257-3282 USA / +1 215-443-8820 Fax Worldwide

Website: www.biodatacorp.com Email: customer.service@biodatacorp.com

APPLICABLE LEGISLATION

Bio/Data Corporation, the Manufacturer, declares that the above-mentioned product meets the provisions of the following European Union legislation:

- Regulation (EU) 2017/746 on In Vitro Diagnostic Medical Devices (IVDR)

Compliance has been demonstrated with the applicable harmonized standards, including:

- EN ISO 13485:2016 + A11:2021 – Medical Devices – Quality Management Systems
- EN ISO 14971:2019 + A11:2021 – Medical Devices – Application of Risk Management to Medical Devices

CONFORMITY STATEMENT

Bio/Data Corporation confirms that the device covered by this Declaration of Conformity is in conformity with Regulation (EU) 2017/746 on In Vitro Diagnostic Medical Devices (IVDR) and, where applicable, with any other relevant Union legislation that provides for the issuing of an EU Declaration of Conformity.

Conformity Assessment Route: Annex II and Annex III of Regulation (EU) 2017/746.

NOTIFIED BODY (NB) INVOLVEMENT

Not applicable. The conformity assessment procedure for Class A (non-sterile) devices is carried out under the sole responsibility of the manufacturer, as such devices are considered to present a low risk to patients.

This device has been self-declared in accordance with Annex II and Annex III of Regulation (EU) 2017/746.

QUALITY ASSURANCE / REGULATORY APPROVAL

Signature: _____



Date: 30 March 2026

William M. Trolio
Director of Quality Assurance & Regulatory Affairs
Bio/Data Corporation
Horsham, PA 19044 USA