



Declaration of Conformity

This European Declaration of Conformity is issued under the sole responsibility of the manufacturer.

MANUFACTURER			
Name of Company	Address	SRN	
Bio/Data Corporation	155 Gibraltar Road, Horsham, PA 19044 U.S.A.	US-MF-00002691	
AUTHORIZED REPRESENTATIVE			
Name of Company	Address	SRN	Telephone/email
mdi Europa GmbH	Langenhagener Str. 71 D-30855 Langenhagen GERMANY	DE-AR-000006218	+49-511-3908 9531 – phone info@mdi-europa.com
PRODUCT IDENTIFICATION			
Product / Trade Name	Product Code / Catalog Number	Basic UDI-DI	
vW Select Ristocetin Cofactor Assay System	106730	++G0561067304W	
Intended Purpose		Photo	
See Instructions for Use		See website www.biodatacorp.com	
The vW Select Ristocetin Cofactor Assay System is optimized to elicit the agglutination of lyophilized platelets in the presence of the patient's Platelet Poor Plasma and Ristocetin.			
IVDR RISK CLASS / COMMON SPECIFICATIONS			
Device Classification		Common Specifications	
Class	A non-sterile	No relevant common specifications have been published yet.	
Rule	5a per Annex VIII of IVDR 2017/746		

155 Gibraltar Road, Horsham, PA 19044 U.S.A.
 Worldwide: (215) 441-4000 U.S.A.: (800) 257-3282
 Fax Worldwide: (215) 443-8820
www.biodatacorp.com e-mail: customer.service@biodatacorp.com
 An ISO 13485 Registered Company

Bio/Data Corporation, the Manufacturer, declares that the above-mentioned product meets the provision of the following EU legislation:

In Vitro Diagnostic Medical Device Regulation (IVDR) (EU) 2017/746

Conformity Statement:

Bio/Data Corporation confirms that the device covered by this declaration is in conformity with the (EU) IVDR 2017/746 Regulation and, if applicable, with any other relevant Union Legislation that provides for the issuing of an EU Declaration of Conformity.

Notified Body (NB) Identification, conformity assessment procedure performed and identification if the certificate or certificates issued:

Not applicable. The conformity assessment procedure for Class A Devices should be carried out, as a general rule, under the sole responsibility of manufacturers, since such devices pose a low risk to patients. (Self-Declaration)

COMPANY REPRESENTATIVE: William M. Trolio

SIGNATURE: _____



TITLE/FUNCTION: Director of Quality Assurance & Regulatory Affairs for Bio/Data Corporation

PLACE: Bio/Data Corporation, Horsham, PA. 19044 U.S.A.

DATE: 1 April 2023